

Print this order form
Fill out the required information
Pack instrument carefully
Include filled form into the package
Send to Vaisala

P.O. Box 26, FI-00421 Helsinki, Finland Tel. +358 9 894 91 Fax. +358 9 8949 2790 Email: helpdesk@vaisala.com www.vaisala.com

Calibration Services for Humidity Indicator HMI41

Please fill in the model of the probe, serial numbers and order code with the selected options:

Model: HM	MI41	Serial no:													
		Serial no (Prob	e):	Co	ode: 223037						Α		Α	Α	
Probe					HMP41										
					HMP42										
					HMP44/44L										
					HMP45										
					HMP46	Ε									
Calibration class					dited calibration		Α								
					oliant calibration		В								
Temperature calibr	ration		No calibration/o	•				Α							
NOTE 1				- ,	+20 and +40 °C			В							
NOTE 2			+17.5, +20.	.0, +22.5, +25	.0 and +27.5 °C			С							
					+40 and +60 °C			D E							
					+40 and +60 °C			Εļ							
			0, +10, +2	20, +30, +40,	+50 and +60 °C			F							
NOTE 3					+60 and +80 °C			G							
Free point 1:	F	ree point 2:	Free point	t 3:				Н							
Humidity calibratio	n				No calibration				Α						
NOTE 4				2, 11, 33,	53 and 75 %RH				В						
			2, 1	1, 33, 53, 75,	85 and 97 %RH				С						
Adjustment					No adjustment					Α					
			Adjuste	d with trimmer	potentiometers					В					
NOTE 5			Adjust	ted with correct	tion coefficients					С					
Calibration method	ds			Using HM	1141 display unit						Α				
Results reported, temperature and humidity %RH and °C											Α				
				C	%RH, °C and °F							В			
Additional reportin	ıg			No add	litional reporting								Α		
Language of the Co	ertificat	е	Certificate in English										•	Α	

NOTE 1: Selected calibration range shall not exceed the measurement range of the probe.

Please note that maximum calibration temperature is +80 °C for HMP46 probes and +60 °C for all other probe types.

NOTE 2: Please note that HMP42 is calibrated only for range from +17,5 to +27,5 $^{\circ}$ C.

NOTE 3: Please fill in the selected free calibration points in range from -20 to +60 °C for HMP41, HMP44, HMP44L and HMP45 and from -40 to +80 °C for HMP46.

NOTE 4: Please note that the humidity calibration is performed using saturated salt solutions at room temperature.

Please note that one-point temperature calibration at +23 °C temperature is included into the humidity calibration.

NOTE 5: The adjustment is done at the indicator and after adjustment the indicator and probe shall be used together.

Requested recalibration stick	er (due-date):			
YES/NO			Interval:	months
Please fill in the end custome	r information for the	Certificate:		
End customer				
l				
Order received and approved				



For other Vaisala locations visit us at: www.vaisala.com